

# Defibrillator Maintenance Record

Defibrillator location: \_\_\_\_\_ Frequency of check: \_\_\_\_\_

Defibrillator make/model: \_\_\_\_\_ Serial number: \_\_\_\_\_

## What to check

### DEFIBRILLATOR

1. Your defibrillator will self-test regularly. Check for any **visible or audible warnings** on the device that something is wrong.
2. Your defibrillator should be in **standby-mode**.
3. Check that your device is **clean** and there is **no damage**.
4. Check the **battery indicator**.
5. Ensure **pads and batteries have not expired**. Reorder if necessary.

### CABINET

1. Check for **damage**.
2. Check that the AED is **easily accessible** in an emergency (key holders? key pad codes?).

Ready for use?	Issues?	Date	Initials

Pads expiry date:    /    /    Reordered    /    /    New expiry    /    /

Batteries expiry date:    /    /    Reordered    /    /    New expiry    /    /

Re-order replacement pads and batteries from [www.safelincs.co.uk](http://www.safelincs.co.uk) or call 0800 612 6537

